



Bath Seats



This information is a general guideline. Please consult a healthcare provider for instructions specific to your situation.

There are many models and styles of bath seats, which are all designed to provide support and safety during bathing. Standard models have a weight limit of 250#; it is important to observe this limit.

Safety Information

- All four leg tips **MUST** be in contact with the shower/tub floor at all times.
- **ALWAYS** inspect the shower chair to ensure that it is properly positioned and stable before using. **DO NOT** use the shower chair if it is wobbly or unstable.
- The shower chair is **NOT** to be used as a transfer bench, transfer device or climbing device.

Adjusting Leg Extension Height

For chairs with rotational legs rotate the leg extension tip counterclockwise to increase leg height or counterclockwise to decrease leg height. Make sure all four legs are adjusted to the same height.

For chairs with snap button legs, press the snap button and slide the leg extension up or down to the desired height. Ensure the snap button fully protrudes through the hole in the leg. Make sure all four legs are adjusted to the same height.